This is a published notice on the Find a Tender service: <a href="https://www.find-tender.service.gov.uk/Notice/000251-2021">https://www.find-tender.service.gov.uk/Notice/000251-2021</a>

Tender

### **Media Autoclave 2020**

Medicines & Healthcare Products Regulatory Agency

F02: Contract notice

Notice identifier: 2021/S 000-000251

Procurement identifier (OCID): ocds-h6vhtk-028841

Published 7 January 2021, 10:00am

## **Section I: Contracting authority**

### I.1) Name and addresses

Medicines & Healthcare Products Regulatory Agency

10 South Colonnade Canary Wharf

**LONDON** 

**E144PU** 

#### Contact

Rene Kiliaan

#### **Email**

rene.kiliaan@nibsc.org

#### **Telephone**

+44 1707641212

### **Country**

**United Kingdom** 

#### **NUTS** code

**UKH - EAST OF ENGLAND** 

#### Internet address(es)

Main address

https://www.gov.uk/government/organisations/medicines-and-healthcare-products-regulatory-agency

### I.3) Communication

The procurement documents are available for unrestricted and full direct access, free of charge, at

https://mhra.bravosolution.co.uk/

Additional information can be obtained from the above-mentioned address

Tenders or requests to participate must be submitted electronically via

https://mhra.bravosolution.co.uk/

### I.4) Type of the contracting authority

National or federal Agency/Office

## I.5) Main activity

Health

## **Section II: Object**

## II.1) Scope of the procurement

### II.1.1) Title

Media Autoclave 2020

Reference number

itt 4596

#### II.1.2) Main CPV code

• 38000000 - Laboratory, optical and precision equipments (excl. glasses)

#### II.1.3) Type of contract

Supplies

#### II.1.4) Short description

A replacement autoclave is required for sterilisation of media and equipment, the current machine (LTE Touchclave R) is old and not fit for purpose due to the time taken to complete each cycle and the capacity / shape of the chamber. Current Cycles required are as shown in the table below (the machine must have capacity for additional cycles to be added)

#### II.1.5) Estimated total value

Value excluding VAT: £50,000

#### II.1.6) Information about lots

This contract is divided into lots: No

### II.2) Description

#### II.2.3) Place of performance

**NUTS** codes

• UKH - EAST OF ENGLAND

Main site or place of performance

National	Institute 1	for Biological	Standards	and Control
Blanche	Lane,			

Potters Bar

South Mimms,

EN63QG

#### II.2.4) Description of the procurement

The new autoclave is required to have:

- Chamber capacity of at least 200 litres
- Overall dimensions, depth no larger than 1200mm (approx. dimensions to be ~1800mm high x 800 mm wide x 1200mm deep)
- A removable shelf
- Armoured ballast (product) probe
- Cooling jacket or similar means of cooling items at the end of the cycle
- Cooling cycle must take no longer than 60 minutes for any load
- A paperless chart recording with the function of being connected via network cable (via CAT 7 data point) to Eurotherm Software
- · Linked to a paper printer as a back up to the paperless system
- Programming must be open code (including autoclave cycles to allow them to be adjusted by competent staff)
- Easily accessible to allow servicing and changing of components to comply with all relevant standards
- All probes must be calibrated to UKAS standards.

#### II.2.5) Award criteria

Price is not the only award criterion and all criteria are stated only in the procurement documents

#### II.2.6) Estimated value

Value excluding VAT: £50,000

#### II.2.7) Duration of the contract, framework agreement or dynamic purchasing system

Duration in days

60

This contract is subject to renewal

No

### II.2.10) Information about variants

Variants will be accepted: Yes

#### II.2.11) Information about options

Options: No

# Section III. Legal, economic, financial and technical information

### III.1) Conditions for participation

#### III.1.2) Economic and financial standing

Selection criteria as stated in the procurement documents

### III.1.3) Technical and professional ability

Selection criteria as stated in the procurement documents

### Section IV. Procedure

## **IV.1) Description**

#### IV.1.1) Type of procedure

Open procedure

#### IV.1.8) Information about the Government Procurement Agreement (GPA)

The procurement is covered by the Government Procurement Agreement: Yes

### IV.2) Administrative information

#### IV.2.2) Time limit for receipt of tenders or requests to participate

Date

7 February 2021

Local time

11:00pm

#### IV.2.4) Languages in which tenders or requests to participate may be submitted

English

#### IV.2.6) Minimum time frame during which the tenderer must maintain the tender

Tender must be valid until: 30 June 2021

#### IV.2.7) Conditions for opening of tenders

Date

8 February 2021

Local time

11:00pm

# **Section VI. Complementary information**

# VI.1) Information about recurrence

This is a recurrent procurement: No

## VI.2) Information about electronic workflows

Electronic ordering will be used

Electronic invoicing will be accepted

Electronic payment will be used

### VI.4) Procedures for review

VI.4.1) Review body

MHRA (Auditors)

10 South Colonnade

London

E14 4PU

Country

**United Kingdom**