

This is a published notice on the Find a Tender service: <https://www.find-tender.service.gov.uk/Notice/000251-2021>

Tender

## **Media Autoclave 2020**

Medicines & Healthcare Products Regulatory Agency

F02: Contract notice

Notice identifier: 2021/S 000-000251

Procurement identifier (OCID): ocds-h6vhtk-028841

Published 7 January 2021, 10:00am

### **Section I: Contracting authority**

#### **I.1) Name and addresses**

Medicines & Healthcare Products Regulatory Agency

10 South Colonnade Canary Wharf

LONDON

E144PU

#### **Contact**

Rene Kiliaan

#### **Email**

[rene.kiliaan@nibsc.org](mailto:rene.kiliaan@nibsc.org)

#### **Telephone**

+44 1707641212

#### **Country**

United Kingdom

**NUTS code**

UKH - EAST OF ENGLAND

**Internet address(es)**

Main address

<https://www.gov.uk/government/organisations/medicines-and-healthcare-products-regulatory-agency>

**I.3) Communication**

The procurement documents are available for unrestricted and full direct access, free of charge, at

<https://mhra.bravosolution.co.uk/>

Additional information can be obtained from the above-mentioned address

Tenders or requests to participate must be submitted electronically via

<https://mhra.bravosolution.co.uk/>

**I.4) Type of the contracting authority**

National or federal Agency/Office

**I.5) Main activity**

Health

---

## **Section II: Object**

### **II.1) Scope of the procurement**

#### **II.1.1) Title**

Media Autoclave 2020

Reference number

itt\_4596

#### **II.1.2) Main CPV code**

- 38000000 - Laboratory, optical and precision equipments (excl. glasses)

#### **II.1.3) Type of contract**

Supplies

#### **II.1.4) Short description**

A replacement autoclave is required for sterilisation of media and equipment, the current machine (LTE Touchclave R) is old and not fit for purpose due to the time taken to complete each cycle and the capacity / shape of the chamber. Current Cycles required are as shown in the table below (the machine must have capacity for additional cycles to be added)

#### **II.1.5) Estimated total value**

Value excluding VAT: £50,000

#### **II.1.6) Information about lots**

This contract is divided into lots: No

### **II.2) Description**

#### **II.2.3) Place of performance**

NUTS codes

- UKH - EAST OF ENGLAND

Main site or place of performance

National Institute for Biological Standards and Control

Blanche Lane,

South Mimms,

Potters Bar

EN6 3QG

#### **II.2.4) Description of the procurement**

The new autoclave is required to have:

- Chamber capacity of at least 200 litres
- Overall dimensions, depth no larger than 1200mm (approx. dimensions to be ~1800mm high x 800 mm wide x 1200mm deep)
- A removable shelf
- Armoured ballast (product) probe
- Cooling jacket or similar means of cooling items at the end of the cycle
- Cooling cycle must take no longer than 60 minutes for any load
- A paperless chart recording with the function of being connected via network cable (via CAT 7 data point) to Eurotherm Software
- Linked to a paper printer as a back up to the paperless system
- Programming must be open code (including autoclave cycles to allow them to be adjusted by competent staff)
- Easily accessible to allow servicing and changing of components to comply with all relevant standards
- All probes must be calibrated to UKAS standards.

#### **II.2.5) Award criteria**

Price is not the only award criterion and all criteria are stated only in the procurement documents

**II.2.6) Estimated value**

Value excluding VAT: £50,000

**II.2.7) Duration of the contract, framework agreement or dynamic purchasing system**

Duration in days

60

This contract is subject to renewal

No

**II.2.10) Information about variants**

Variants will be accepted: Yes

**II.2.11) Information about options**

Options: No

---

**Section III. Legal, economic, financial and technical information**

**III.1) Conditions for participation**

**III.1.2) Economic and financial standing**

Selection criteria as stated in the procurement documents

**III.1.3) Technical and professional ability**

Selection criteria as stated in the procurement documents

---

## **Section IV. Procedure**

### **IV.1) Description**

#### **IV.1.1) Type of procedure**

Open procedure

#### **IV.1.8) Information about the Government Procurement Agreement (GPA)**

The procurement is covered by the Government Procurement Agreement: Yes

### **IV.2) Administrative information**

#### **IV.2.2) Time limit for receipt of tenders or requests to participate**

Date

7 February 2021

Local time

11:00pm

#### **IV.2.4) Languages in which tenders or requests to participate may be submitted**

English

#### **IV.2.6) Minimum time frame during which the tenderer must maintain the tender**

Tender must be valid until: 30 June 2021

#### **IV.2.7) Conditions for opening of tenders**

Date

8 February 2021

Local time

11:00pm

---

## **Section VI. Complementary information**

### **VI.1) Information about recurrence**

This is a recurrent procurement: No

### **VI.2) Information about electronic workflows**

Electronic ordering will be used

Electronic invoicing will be accepted

Electronic payment will be used

### **VI.4) Procedures for review**

#### **VI.4.1) Review body**

MHRA (Auditors)

10 South Colonnade

London

E14 4PU

Country

United Kingdom