

This is a published notice on the Find a Tender service: <https://www.find-tender.service.gov.uk/Notice/000251-2021>

Tender

Media Autoclave 2020

Medicines & Healthcare Products Regulatory Agency

F02: Contract notice

Notice identifier: 2021/S 000-000251

Procurement identifier (OCID): ocds-h6vhtk-028841

Published 7 January 2021, 10:00am

Section I: Contracting authority

I.1) Name and addresses

Medicines & Healthcare Products Regulatory Agency

10 South Colonnade Canary Wharf

LONDON

E144PU

Contact

Rene Kiliaan

Email

rene.kiliaan@nibsc.org

Telephone

+44 1707641212

Country

United Kingdom

NUTS code

UKH - EAST OF ENGLAND

Internet address(es)

Main address

<https://www.gov.uk/government/organisations/medicines-and-healthcare-products-regulatory-agency>

I.3) Communication

The procurement documents are available for unrestricted and full direct access, free of charge, at

<https://mhra.bravosolution.co.uk/>

Additional information can be obtained from the above-mentioned address

Tenders or requests to participate must be submitted electronically via

<https://mhra.bravosolution.co.uk/>

I.4) Type of the contracting authority

National or federal Agency/Office

I.5) Main activity

Health

Section II: Object

II.1) Scope of the procurement

II.1.1) Title

Media Autoclave 2020

Reference number

itt_4596

II.1.2) Main CPV code

- 38000000 - Laboratory, optical and precision equipments (excl. glasses)

II.1.3) Type of contract

Supplies

II.1.4) Short description

A replacement autoclave is required for sterilisation of media and equipment, the current machine (LTE Touchclave R) is old and not fit for purpose due to the time taken to complete each cycle and the capacity / shape of the chamber. Current Cycles required are as shown in the table below (the machine must have capacity for additional cycles to be added)

II.1.5) Estimated total value

Value excluding VAT: £50,000

II.1.6) Information about lots

This contract is divided into lots: No

II.2) Description

II.2.3) Place of performance

NUTS codes

- UKH - EAST OF ENGLAND

Main site or place of performance

National Institute for Biological Standards and Control

Blanche Lane,

South Mimms,

Potters Bar

EN6 3QG

II.2.4) Description of the procurement

The new autoclave is required to have:

- Chamber capacity of at least 200 litres
- Overall dimensions, depth no larger than 1200mm (approx. dimensions to be ~1800mm high x 800 mm wide x 1200mm deep)
- A removable shelf
- Armoured ballast (product) probe
- Cooling jacket or similar means of cooling items at the end of the cycle
- Cooling cycle must take no longer than 60 minutes for any load
- A paperless chart recording with the function of being connected via network cable (via CAT 7 data point) to Eurotherm Software
- Linked to a paper printer as a back up to the paperless system
- Programming must be open code (including autoclave cycles to allow them to be adjusted by competent staff)
- Easily accessible to allow servicing and changing of components to comply with all relevant standards
- All probes must be calibrated to UKAS standards.

II.2.5) Award criteria

Price is not the only award criterion and all criteria are stated only in the procurement documents

II.2.6) Estimated value

Value excluding VAT: £50,000

II.2.7) Duration of the contract, framework agreement or dynamic purchasing system

Duration in days

60

This contract is subject to renewal

No

II.2.10) Information about variants

Variants will be accepted: Yes

II.2.11) Information about options

Options: No

Section III. Legal, economic, financial and technical information

III.1) Conditions for participation

III.1.2) Economic and financial standing

Selection criteria as stated in the procurement documents

III.1.3) Technical and professional ability

Selection criteria as stated in the procurement documents

Section IV. Procedure

IV.1) Description

IV.1.1) Type of procedure

Open procedure

IV.1.8) Information about the Government Procurement Agreement (GPA)

The procurement is covered by the Government Procurement Agreement: Yes

IV.2) Administrative information

IV.2.2) Time limit for receipt of tenders or requests to participate

Date

7 February 2021

Local time

11:00pm

IV.2.4) Languages in which tenders or requests to participate may be submitted

English

IV.2.6) Minimum time frame during which the tenderer must maintain the tender

Tender must be valid until: 30 June 2021

IV.2.7) Conditions for opening of tenders

Date

8 February 2021

Local time

11:00pm

Section VI. Complementary information

VI.1) Information about recurrence

This is a recurrent procurement: No

VI.2) Information about electronic workflows

Electronic ordering will be used

Electronic invoicing will be accepted

Electronic payment will be used

VI.4) Procedures for review

VI.4.1) Review body

MHRA (Auditors)

10 South Colonnade

London

E14 4PU

Country

United Kingdom