

This is a published notice on the Find a Tender service: <https://www.find-tender.service.gov.uk/Notice/001633-2024>

Planning

Supply of Demineralised Bone Matrix Syringe

NHS Blood and Transplant

F01: Prior information notice

Prior information only

Notice identifier: 2024/S 000-001633

Procurement identifier (OCID): ocds-h6vhtk-042df2

Published 17 January 2024, 2:41pm

Section I: Contracting authority

I.1) Name and addresses

NHS Blood and Transplant

500 North Bristol Park

Bristol

BS34 7QH

Contact

Sahithi Bandi

Email

sahithi.bandi@nhsbt.nhs.uk

Country

United Kingdom

Region code

UKK11 - Bristol, City of

Internet address(es)

Main address

<https://www.nhsbt.nhs.uk/>

Buyer's address

<https://www.nhsbt.nhs.uk/>

I.3) Communication

Additional information can be obtained from the above-mentioned address

Electronic communication requires the use of tools and devices that are not generally available. Unrestricted and full direct access to these tools and devices is possible, free of charge, at

<https://health-family.force.com/s/Welcome>

I.4) Type of the contracting authority

Body governed by public law

I.5) Main activity

Health

Section II: Object

II.1) Scope of the procurement

II.1.1) Title

Supply of Demineralised Bone Matrix Syringe

II.1.2) Main CPV code

- 33000000 - Medical equipments, pharmaceuticals and personal care products

II.1.3) Type of contract

Supplies

II.1.4) Short description

Supply of Demineralised Bone Matrix Syringe

NHS Blood & Transplant (NHSBT) intends to undertake a procurement tender exercise for syringe to package Demineralised Bone Matrix Paste and Putty

NHSBT now wishes to commence a formal procurement process and is seeking to engage with suppliers available in the marketplace. This will aid identification of opportunities and gauge market interest in this field.

II.1.5) Estimated total value

Value excluding VAT: £200,000

II.1.6) Information about lots

This contract is divided into lots: No

II.2) Description

II.2.2) Additional CPV code(s)

- 33000000 - Medical equipments, pharmaceuticals and personal care products

II.2.3) Place of performance

NUTS codes

- UKK11 - Bristol, City of

II.2.4) Description of the procurement

Supply of Demineralised Bone Matrix Syringe

NHS Blood & Transplant (NHSBT) intends to undertake a procurement tender exercise for syringe to package Demineralised Bone Matrix Paste and Putty

NHSBT now wishes to commence a formal procurement process and is seeking to engage with suppliers available in the marketplace. This will aid identification of opportunities and gauge market interest in this field.

Requirements

NHSBT TES process cortical bone into DBM Paste and Putty grafts. The cortical bone is donated from deceased donors. The syringe must be suitable for containing DBM paste and putty grafts and allow dispense of this product during surgery.

Technical Requirements

- If provided sterile, syringe must be packaged and sterilised (not by gamma-irradiation) with a visible confirmation of sterilisation and expiry date for packaging on each pack.
- If provided sterile, plunger must not be inside barrel of syringe prior to initial sterilisation if ETO sterilisation is used. This is to ensure all surfaces of syringe are sterile.
- Must be suitable for sterilisation by gamma-irradiation (25-40kGy) with the graft inside.
- Syringe must be of a suitable size to hold the DBM Paste and Putty in 1cc, 5cc, and 10cc sizes and have the plunger placed inside. The syringe must be easily handheld for use within theatre.
- Must be an open bore syringe supplied with a dead-end cap.
- Must be able to be stored at -80°C in a freezer and on dry ice.
- Syringe must be made of polycarbonate material and must not be opaque.
- Expiry date of each batch must be a minimum of 1 year from date of delivery.
- The package will be CE marked for the purpose of tissue storage.

Environmental/Sustainability Requirements

- NHSBT requires suppliers to take responsibility for their impact on the Environment and where relevant society. Suppliers are required to demonstrate an Environmental Management System e.g. ISO14001 or an equivalent EMS. NHSBT may require the auditing of this system. Where an EMS does not exist suppliers will be required to demonstrate adequate management of their impact and where suitable may be required to develop an EMS system.
- Suppliers must comply with all legislation governing Environmental obligations. Failure to do so will result in exclusion from tendering.

Stock Holding and Business Continuity – Quality/EP Critical Only

- The Contractor must hold a minimum un-reserved buffer stock of one month's usage (based on NHSBT's normal volume usage), unless otherwise agreed by NHSBT.
- Two different batches should be available for delivery to any of the Authority's warehouses or centres within 24 hours, the Contractor should ensure their stock is rotated as usual as no obsolete stock will be NHSBT's responsibility.
- In addition to these minimum requirements extra stock may need to be held in the supply chain. In such a case, the Contractor and NHSBT will from time to time agree appropriate stock holding requirements, and these will be confirmed in writing by NHSBT.
- Suppliers may be required to confirm stock levels, batch numbers and expiry dates upon request from NHSBT.
- As a supplier of critical consumables and/or services to NHSBT, the supplier is required to have a documented Business Continuity plan covering, at least, all aspects and locations of their business which impact the supply of the goods or services which are supplied to NHSBT. Wherever possible any such plan should have been constructed to comply with the relevant British Standard for Business Continuity (BS 25999 Parts 1&2). In the future NHSBT may seek evidence of independent certification to this standard.
- Any supplier Business Continuity plan as above should be regularly reviewed, updated, and tested or exercised at least every two years.

It is NHSBT's intention to hold pre-procurement supplier engagement meetings the week of 12th February 2024 to 23rd February 2024. the meetings will be held remotely; the dates are to be confirmed. To enable NHSBT to finalise arrangements (e.g. send meeting details and agree time slots) for the meeting's, interested organisations are required to send an expression of interest via email to Sahithi Bandi - sahithi.bandi@nhsbt.nhs.uk no later than 17:00 hrs (GMT) on 09/02/2024. This should include a list of your potential attendees and a brief overview (or a website link) of the Goods / Services which might be offered.

No business whatsoever is guaranteed under any resulting framework agreement or contract, and there is no guarantee that any framework agreement or contract will be put in place in relation to this notice. No compensation etc. will be paid if a tender or resulting framework agreement is withdrawn for any reason. Bidders should take part in this process only on the basis that they fully understand and accept this position.

It should be noted that this procurement exercise may allow other Health authorities to use any agreements awarded.

II.3) Estimated date of publication of contract notice

17 January 2024

Section IV. Procedure

IV.1) Description

IV.1.8) Information about the Government Procurement Agreement (GPA)

The procurement is covered by the Government Procurement Agreement: Yes