

This is a published notice on the Find a Tender service: <https://www.find-tender.service.gov.uk/Notice/021056-2024>

Award

VEAT_129092-External Beam Gamma Teletherapy System

National Physical Laboratory

F15: Voluntary ex ante transparency notice

Notice identifier: 2024/S 000-021056

Procurement identifier (OCID): ocds-h6vhtk-047c3b

Published 10 July 2024, 12:54pm

Section I: Contracting authority/entity

I.1) Name and addresses

National Physical Laboratory

Hampton Road

Teddington

TW11 0LW

Email

gary.phillips@npl.co.uk

Telephone

+44 2089773222

Country

United Kingdom

Region code

UK - United Kingdom

Internet address(es)

Main address

www.npl.co.uk

I.4) Type of the contracting authority

Body governed by public law

I.5) Main activity

Other activity

Research

Section II: Object

II.1) Scope of the procurement

II.1.1) Title

VEAT_129092-External Beam Gamma Teletherapy System

II.1.2) Main CPV code

- 38000000 - Laboratory, optical and precision equipments (excl. glasses)

II.1.3) Type of contract

Supplies

II.1.4) Short description

The NPL requirement is for an External Beam Gamma Teletherapy System using Cobalt-60 source. This is to replace our current unit which is and failing due to the obsolescence of the device making replacement parts difficult to source. The device must be able to move the

Cobalt-60 source to its exposed position consistently within 0.1 seconds as well as the position being reproducible within 0.1mm.

II.1.6) Information about lots

This contract is divided into lots: No

II.1.7) Total value of the procurement (excluding VAT)

Value excluding VAT: 378,150 USD

II.2) Description

II.2.3) Place of performance

NUTS codes

- UK - United Kingdom

Main site or place of performance

Teddington, TW11 0LW

II.2.4) Description of the procurement

The NPL requirement is for an External Beam Gamma Teletherapy System using Cobalt-60 source. This is to replace our current unit which is and failing due to the obsolescence of the device making replacement parts difficult to source. The device must be able to move the Cobalt-60 source to its exposed position consistently within 0.1 seconds as well as the position being reproducible within 0.1mm.

II.2.11) Information about options

Options: No

II.2.13) Information about European Union Funds

The procurement is related to a project and/or programme financed by European Union funds: No

Section IV. Procedure

IV.1) Description

IV.1.1) Type of procedure

Negotiated without a prior call for competition

- The products involved are manufactured purely for the purpose of research, experiment, study or development
- The works, supplies or services can be provided only by a particular economic operator for the following reason:
 - absence of competition for technical reasons

Explanation:

After a thorough look at the market only one supplier capable of achieving the required specification was identified.

This supplier supplied our current Cobalt-60 source and their irradiators are designed to take this source type.

IV.1.8) Information about the Government Procurement Agreement (GPA)

The procurement is covered by the Government Procurement Agreement: Yes

Section V. Award of contract/concession

Contract No

129092

Title

External Beam Gamma Teletherapy System

A contract/lot is awarded: Yes

V.2) Award of contract/concession

V.2.1) Date of conclusion of the contract

10 July 2024

V.2.2) Information about tenders

The contract has been awarded to a group of economic operators: No

V.2.3) Name and address of the contractor/concessionaire

Best Theratronics

413 March Road

Ottawa

ON K2K OE4

Country

Canada

NUTS code

- CA - Canada

The contractor/concessionaire is an SME

No

V.2.4) Information on value of contract/lot/concession (excluding VAT)

Total value of the contract/lot/concession: 378,150 USD

V.2.5) Information about subcontracting

The contract/lot/concession is likely to be subcontracted

Section VI. Complementary information

VI.3) Additional information

After a thorough look at the market only one supplier capable of achieving the required specification was identified.

This supplier supplied our current Cobalt-60 source and their irradiators are designed to take this source type.

VI.4) Procedures for review

VI.4.1) Review body

NPL Management Limited

Teddington

TW11 0LW

Country

United Kingdom