This is a published notice on the Find a Tender service: https://www.find-tender.service.gov.uk/Notice/023092-2025

Award

Dry heat sterilizer for animal bedding and cages

University Of Edinburgh

F15: Voluntary ex ante transparency notice

Notice identifier: 2025/S 000-023092

Procurement identifier (OCID): ocds-h6vhtk-05184e

Published 20 May 2025, 11:03am

Section I: Contracting authority/entity

I.1) Name and addresses

University Of Edinburgh

Charles Stewart House, 9-16 Chambers Street

Edinburgh

EH1 1HT

Email

jpike2@ed.ac.uk

Telephone

+44 1316502759

Country

United Kingdom

NUTS code

UKM75 - Edinburgh, City of

Internet address(es)

Main address

http://www.ed.ac.uk

Buyer's address

https://www.publiccontractsscotland.gov.uk/search/Search_AuthProfile.aspx?ID=AA0010 7

I.4) Type of the contracting authority

Body governed by public law

I.5) Main activity

Education

Section II: Object

II.1) Scope of the procurement

II.1.1) Title

Dry heat sterilizer for animal bedding and cages

Reference number

NCA31110

II.1.2) Main CPV code

• 42000000 - Industrial machinery

II.1.3) Type of contract

Supplies

II.1.4) Short description

The University provides accommodation for a large volume of animals in the course of it's

day to day work. Regular sterilization of bedding and cages is critical to ensure the health and wellbeing of animals and to comply with bio-security controls set by the government and by accepted best-practice (such as avoiding the propagation of infectious diseases).

In recent months an incumbent steam-based sterilizer system has failed on a regular basis, placing these bio-security and welfare controls at risk. Steam based machines are time consuming to build and maintain; in order to avoid an operational incident and provide biosecurity assurance as soon as possible the University has decided it should urgently procure an immediate solution, ahead of making long term plans for the replacement/upgrade of such machines. It is believed that a new type of device — a dryheat based steriliser — would be the most appropriate solution, as in principle this is a simpler device which could be delivered and built faster than a steam counterpart.

On the basis of;

- Urgent circumstances which the University could not reasonably foresee or control
- It is believed there may only be one supplier on the market who has an established dryheat proposition with a prior service record encompassing a reasonable period of time

The University is notifying the market it plans to undertake a Negotiated Procedure Without Prior Call for Competition.

The University will observe a ten (10) day Stand-Still in line with the Public Contract Regulations (Scotland) 2015. Any suppliers wishing to challenge this procedure should do so with the individual stated on the notice, during this Stand-Still.

Value is an estimate consisting of GBP 380k expected cost with GBP 25k contingency.

II.1.6) Information about lots

This contract is divided into lots: No

II.1.7) Total value of the procurement (excluding VAT)

Value excluding VAT: £405,000

II.2) Description

II.2.3) Place of performance

NUTS codes

• UKM75 - Edinburgh, City of

Main site or place of performance

City of Edinburgh

II.2.4) Description of the procurement

Technical requirements

A dry heat sterilizer is required for the routine sterilization of assembled animal cages, bedding and other thermostable items involved in laboratory animal science.

General characteristics:

- -Electrically powered heating, either 240V (2-phase) or 400V (3-phase)
- -Operating temperature of 140 deg C for routine use, no higher than 210°C at any time
- -Air circulation required to ensure even temperature distribution throughout static holding period
- -Intake air must be filtered to