

This is a published notice on the Find a Tender service: <https://www.find-tender.service.gov.uk/Notice/026279-2023>

Planning

Antimicrobial Susceptibility Testing (AST) system

The Common Services Agency (more commonly known as NHS National Services Scotland) (NSS)

F01: Prior information notice

Prior information only

Notice identifier: 2023/S 000-026279

Procurement identifier (OCID): ocds-h6vhtk-03fa51

Published 6 September 2023, 9:50am

Section I: Contracting authority

I.1) Name and addresses

The Common Services Agency (more commonly known as NHS National Services Scotland) (NSS)

Gyle Square (NSS Head Office), 1 South Gyle Crescent

Edinburgh

EH12 9EB

Contact

Shaun Kean

Email

shaun.kean@nhs.scot

Telephone

+44 1698794400

Country

United Kingdom

NUTS code

UKM - Scotland

Internet address(es)

Main address

<http://www.nss.nhs.scot/browse/procurement-and-logistics>

Buyer's address

https://www.publiccontractsscotland.gov.uk/search/Search_AuthProfile.aspx?ID=AA11883

I.2) Information about joint procurement

The contract is awarded by a central purchasing body

I.3) Communication

Additional information can be obtained from the above-mentioned address

I.4) Type of the contracting authority

Body governed by public law

I.5) Main activity

Health

Section II: Object

II.1) Scope of the procurement

II.1.1) Title

Antimicrobial Susceptibility Testing (AST) system

Reference number

NP

II.1.2) Main CPV code

- 33100000 - Medical equipments

II.1.3) Type of contract

Supplies

II.1.4) Short description

NHS Scotland requires a standard Antimicrobial Susceptibility Testing (AST) system to be used across all 15 NHS Health Boards which provide a Clinical Microbiology service: from multi-disciplinary laboratories in remote / rural Boards to large laboratories providing specialist services in tertiary referral Boards.

The AST solution must be (semi-) automated and up-scalable so that it is practical for both small and large laboratories and it must have a proven real-world performance record.

Please note interest by 31st October 2023

II.1.6) Information about lots

This contract is divided into lots: No

II.2) Description

II.2.3) Place of performance

NUTS codes

- UKM - Scotland

Main site or place of performance

Health Boards across Scotland

II.2.4) Description of the procurement

NHS Scotland requires a standard Antimicrobial Susceptibility Testing (AST) system to be used across all 15 NHS Health Boards which provide a Clinical Microbiology service: from multi-disciplinary laboratories in remote / rural Boards to large laboratories providing specialist services in tertiary referral Boards.

The AST solution must be (semi-) automated and up-scalable so that it is practical for both small and large laboratories and it must have a proven real-world performance record.

The national AST system must deliver panel testing for both bacteria and fungi. Suppliers are requested to provide details of the antibiotics, organisms and “drug-bug” combinations which may be tested on their system as well as those which cannot be tested on their system.

It is essential that NHS Scotland-bespoke testing panels are available and that the system is fully EUCAST compliant, in line with Scottish AMR surveillance requirements. Annual EUCAST updates must be available on the system within six months of publication, including breakpoints and interpretive criteria. In the absence of specific EUCAST “drug-bug” testing and interpretive criteria, alternatives stipulated by the SMVN (e.g. CLSI) must be incorporated into the testing system.

If the supplier’s system is capable of delivering “rapid AST (RAST)”, details should be provided.

A bi-directional interface between the AST system and each Board’s Laboratory Information Management System (LIMS) is essential and implementation should be led by the AST system supplier.

The national AST solution must routinely report minimum inhibitory concentration (MIC) data as well interpreted MICs via system software to routinely report antibiograms and ensure unusual phenotypes are highlighted. The ability to search and produce management reports on AST data would be advantageous and details should be provided.

AST equipment is expected to have a lifespan of a minimum of 7 to 10 years.

Suppliers are asked to submit details of [a] their standard (semi-) automated AST equipment currently available for all 15 NHS Scotland Boards that deliver an AST service, [b] details of standard testing panels currently available, [c] interpretive criteria, [d]

reportable patient results and [e] epidemiological packages, if available. Prerequisites and dependence on any additional or ancillary equipment / systems should be highlighted.

Suppliers are asked to describe the system(s) available to fulfil this national requirement and also provide a (maximum) 2 page summary detailing the service that would be provided.

Indicative numbers of ASTs carried out per year in NHS Scotland laboratories (based on data from 2018 – 2022) are:

AST tests per year

7,000 1 board

II.3) Estimated date of publication of contract notice

12 February 2024

Section IV. Procedure

IV.1) Description

IV.1.8) Information about the Government Procurement Agreement (GPA)

The procurement is covered by the Government Procurement Agreement: Yes

Section VI. Complementary information

VI.3) Additional information

Suppliers who believe they have a solution to the requirement outlined in this notice, should submit a brief summary of their proposed equipment solution along with technical manuals describing the equipment and software to be utilised.

Information on where these solutions are currently deployed should also be provided if available

NOTE: To register your interest in this notice and obtain any additional information please visit the Public Contracts Scotland Web Site at

https://www.publiccontractsscotland.gov.uk/Search/Search_Switch.aspx?ID=743858.

(SC Ref:743858)

