

This is a published notice on the Find a Tender service: <https://www.find-tender.service.gov.uk/Notice/032854-2022>

Not applicable

In Vivo Imaging System

London School of Hygiene & Tropical Medicine

F14: Notice for changes or additional information

Notice identifier: 2022/S 000-032854

Procurement identifier (OCID): ocds-h6vhtk-037cfb

Published 18 November 2022, 4:22pm

Section I: Contracting authority/entity

I.1) Name and addresses

London School of Hygiene & Tropical Medicine

Keppel Street,

London

WC1E 7HT

Contact

Paul Pester

Email

paul.pesther@lshtm.ac.uk

Telephone

+44 02079272471

Country

United Kingdom

Region code

UKI31 - Camden and City of London

National registration number

RC000330

Internet address(es)

Main address

<http://www.lshtm.ac.uk>

Buyer's address

<https://uk.eu-supply.com/ctm/Company/CompanyInformation/Index/104519>

Section II: Object

II.1) Scope of the procurement

II.1.1) Title

In Vivo Imaging System

Reference number

LSHTM-2022-50

II.1.2) Main CPV code

- 33110000 - Imaging equipment for medical, dental and veterinary use

II.1.3) Type of contract

Supplies

II.1.4) Short description

Purchase of a ultra-high resolution optical imaging system including highly sensitive fluorescence and bioluminescence in vivo imaging, 3D Fluorescence and tomography and X-ray computed Tomography (CT) to be delivered, installed and commissioned in LSHTM Keppel Street, London, WC1E 7HT by March 2023.

Section VI. Complementary information

VI.6) Original notice reference

Notice number: [2022/S 000-030427](#)

Section VII. Changes

VII.1.2) Text to be corrected in the original notice

Section number

11.2.4

Instead of

Text

The in vivo imaging system must be capable of the following;

- Highly sensitive 2D fluorescence and bioluminescence in vivo imaging. 4 magnification settings, the ability to image 10 mice simultaneously (with temperature and anaesthesia control), and computer-regulated narrow band-pass excitation filters covering the range of 300-900 nm.
- 3D-tomographic reconstruction for both bioluminescence and fluorescence using real CT data to create a non-uniform light absorption/scattering map. The 3D reconstruction to correct for non-uniform light absorption/scattering by different organs/tissue allowing for non-invasive and much more precise assessment of heart, GI tract and CNS infections.
- The X-ray CT facility must be able to combines scan times down to 5 seconds for total body mouse imaging, ultra-low X-ray dosage (down to 2 mGy), and an ability to image 4 mice simultaneously.

Read

Text

The in vivo imaging system should be capable of the following:

- Highly sensitive 2D fluorescence and bioluminescence in vivo imaging. 4 Fields of View, the ability to image optically 10 mice simultaneously (with temperature and anaesthesia control), using narrow band-pass excitation filters covering the range of (420-740nm).
- 3D-tomographic reconstruction for both bioluminescence and fluorescence using real CT data to create a co registered image. Retrospective CT gating to allow precise assessment of cardiac and respiratory gating. The system should also be able to carry out GI tract and CNS infection imaging.
- The X-ray CT facility must be able to carry out total body mouse imaging, using low X-ray dosage.

Section number

11.2.10

Instead of

Text

No

Read

Text

Yes