

This is a published notice on the Find a Tender service: <https://www.find-tender.service.gov.uk/Notice/069601-2025>

Planning

## Purchase of Portable Human MRI System

Cardiff University

UK2: Preliminary market engagement notice - Procurement Act 2023 - [view information about notice types](#)

Notice identifier: 2025/S 000-069601

Procurement identifier (OCID): ocds-h6vhtk-05d6ca ([view related notices](#))

Published 30 October 2025, 9:33am

### Scope

### Reference

CU.2082.AR

### Description

Cardiff University is currently exploring the acquisition of a Portable Human MRI System. Our comprehensive market research indicates that there is only one supplier available for this specific technology. In light of this, we are issuing this Preliminary market engagement notice to invite expressions of interest from any suppliers who believe they can also meet our requirements.

We encourage interested parties to submit their expressions of interest by emailing [rogalaa@cardiff.ac.uk](mailto:rogalaa@cardiff.ac.uk) no later than Thursday, November 13, 2025. In the event that we do not receive any responses, the contract will be awarded directly in accordance with Regulation 41.

Requirements:

Portable human MRI system operating at very low field strength and suitable for imaging the brain in vivo.

The system should use permanent magnets and must be modular in hardware (i.e. individual parts can be replaced with customised components, but it should come complete of gradients and an RF head coil).

The code for running MRI sequences should be open source and freely available (i.e. users must be able to program any acquisition system, not just modify parameters of available ones).

No external cooling should be required and it should be powered by a standard single-phase electricity supply.

K-space and complex imaging data need to be exportable.

- Target anatomy Head
- Field strength (B0): between 40 and 80 mT.
- Field homogeneity (over 20 cm DSV): Less than 1000 ppm.
- Max gradient strength on each channel simultaneously: Greater than 25 mT/m.
- Max gradient strength on each channel simultaneously
- Number of Rx Channels 8 or more
- Hardware Modular and customisable
- Software (acquisition) Open source
- Software (reconstruction) Open source

### **Total value (estimated)**

- £400,000 excluding VAT
- £400,000 including VAT

Above the relevant threshold

### **Contract dates (estimated)**

- 1 December 2025 to 30 November 2026
- 1 year

### **Main procurement category**

Goods

### **CPV classifications**

- 33100000 - Medical equipments

### **Contract locations**

- UK - United Kingdom

---

## **Engagement**

### **Engagement deadline**

13 November 2025

### **Engagement process description**

We encourage interested parties to submit their expressions of interest by emailing [rogalaa@cardiff.ac.uk](mailto:rogalaa@cardiff.ac.uk) no later than Thursday, November 13, 2025.

---

## Participation

### Particular suitability

- Small and medium-sized enterprises (SME)
- Voluntary, community and social enterprises (VCSE)

---

## Contracting authority

### Cardiff University

- Public Procurement Organisation Number: PHVT-4343-HWYR

Procurement Services

Cardiff

CF24 4HQ

United Kingdom

Email: [rogalaa@cardiff.ac.uk](mailto:rogalaa@cardiff.ac.uk)

Website: <http://www.cardiff.ac.uk/business/why-work-with-us/for-suppliers>

Region: UKL22 - Cardiff and Vale of Glamorgan

Organisation type: Public authority - sub-central government

Devolved regulations that apply: Wales